Microalbumin

METHOD

Immunoturbidimetric

TECHNICAL INFORMATION

Storage Temperature:	2-8°C
Reconstituted Stability:	Ready-to-use Liquid
Linearity:	0.5-30.0 mg/dL
Wavelength:	340 nm
Expected Values:	Less than 2.1 mg/dL (24 hours)

ORDERING INFORMATION

CATALOG #		SIZE
M7562-100	Buffer	4 x 20 ml
	Antiserum	2 x 10 ml

Calibrator Set Required

RELATED PRODUCTS

M7562-CAL	Microalbumin Calibrator Set
M7562-CTL	Microalbumin Control Set